



CALCUTTA SCHOOL OF TROPICAL MEDICINE

GOVERNMENT OF WEST BENGAL

108 Chittaranjan Avenue, Kolkata - 700073

Phone : 033 2212 3695/96/97 Fax : 033 2212 3698

Website : stmkkolkata.org

Memo No. STM/STORE/QUOTE/044/2023-24 QUOTATION NOTICE

Date:04.07.2023

Sealed QUOTATIONS are invited from the Reputed Company/Direct Importers/Authorized distributor firms for supply of items as mentioned below at the Department of CCU, **School of Tropical Medicine, Kolkata.**

The QUOTATION is to be submitted in the drop box which will be kept at the office chamber of the undersigned on all working days from 05.07.2023

The last date of submission of QUOTATION is **12.07.2023 up to 12.00 pm** and the same will be opened on the same date in presence of interested bidder, if any. The bidders will submit the QUOTATION in their own official pad along with the following essential documents :

- Valid Trade License / Enlistment
- PAN card of the Bidder /Bidder Company
- GST Registration certificate
- IT Return of the last financial year

The financial quotation Should be quoted as per the following table only :-

Item	Unit Price	GST	Total price
------	------------	-----	-------------

Physical verification of samples and or physical demonstration may be done at the discretion of the hospital authority by a team of expert to adjudge the applicability or suitability of the product or products in the functional requirement of the patients since the items are life savings and the technical evaluation will be done accordingly in close observation of the product or products, technical brochures along with Test certificate produced and physical sampling thereof.

Quotation Drop box STM Main Building (2nd floor) MSVP Office .Monday to Friday At 11am To 4pm and Saturday at 11am to 2pm except Sundays and holidays.

No hand written document accepted only printed documents submitted.

Sl. No	Items
1.	Continuous Enteral Feeding Pump compatible with feeding bag DEHP/LATEX free with Extra medication port & EN Fit connector. (Remarks: 30 unit of feeding bag should be supply with feeding pump) (Specification Attached)

Raw 4/7/2023
Raw 4/7/2023

Medical Superintendent cum Vice Principal,
School of Tropical Medicine, Kolkata

MSVP
School of Tropical Medicine
Kolkata

Sl. No.	Specification - Enteral Feeding Pump
1	Should work on linear peristaltic mechanism
2	Flow rate range should be 1 - 600 mL/h. 1 mL/h increment from 1 to 100 mL/h, 5 mL/h from 100 to 600 mL/h.
3	Should work on closed or dedicated giving sets
4	All Closed or dedicated giving sets should be DHEP-free & Latex-Free
5	Feeding volume range should be 1 - 5000 ml
6	Should operate on continuous feeding mode
7	Automatic, semi-automatic and manual priming facility should be available
8	Should have cumulative feeding counter from 0.001L - 99.999L
9	Should have following data event logs: Feeding history - 250 events Alarm history - 250 events
10	Should have night mode to decrease screen and power LED brightness
11	Should have keypad locking facility
12	Should have air in line detection feature
13	Should have occlusion detection feature
14	The device should have the following alarm: Target Volume Reached, Door Open, wrong set installation, downstream occlusion, upstream occlusion, empty bag/ air in line, empty battery and technical error
15	Should have LCD display, size should be equal or more than 40 mm x 59 mm
16	Should have Swinglock clamp for versatile clamp and horizontal clamp that allows the fixation on a rail or on a pole
17	Should have rechargeable NiMh battery, remaining battery life indicator available on the display. Battery Life (when fully charged): 25 h at 125 mL/h
18	Battery charging time should not be more than 6 hrs
19	Should have EN/IEC 60601-1:2006 and EN/IEC 60601-1-2:2007 compliance standard
20	Medical Device Directive - CE marking in compliance with the Council Directive 93/42/EEC
21	Electrical Compliance - Protection against leakage current: Defibrillation-proof type CF and Protection against electric shocks: class II
22	Should be IP34 certified for waterproofness
23	All Closed or dedicated giving sets should be DHEP-free & Latex-Free
24	30 units of feeding bag should be supplied with feeding pump

Debayoti
21/5

DR. DEBAJYOTI MAJUMDAR
Clinical Tutor
DEPT. OF TROPICAL MEDICINE
SCHOOL OF TROPICAL MEDICINE
Kolkata-700073